

DIVISION OF AIR QUALITY
Ambient Monitoring Section
 217 West Jones Street / 1641 Mail Service Center
 Raleigh, NC 27699-1641

7/26/2022

QA BULLETIN NUMBER DAQ-17-012.2

TO: Marshall Cannon, RRO Operator DocuSigned by: [Signature] 7/19/2022
5A799BD9BE664DE...

Steve Helms, RRO Operator DocuSigned by: 8/1/2022
[Signature]

Mark Yirka, ECB Electronics Technician DocuSigned by: 8/1/2022
[Signature]

THROUGH: Patrick Bulter, AMS Chief DocuSigned by: 7/18/2022
[Signature]

Joette Steger, PPB Supervisor DocuSigned by: 7/18/2022
[Signature]

Derrick House, ECB Supervisor DocuSigned by: 8/1/2022
[Signature]

Taylor Hartsfiend, RRO Air Quality Supervisor DocuSigned by: 7/18/2022
[Signature]

Tim Skelding, RRO Monitoring Coordinator DocuSigned by: 7/19/2022
[Signature]

FROM: Kay Roberts, PPB Chemist DocuSigned by: 7/18/2022
[Signature]

SUBJECT: **Zero Criteria Warning and Control Limits for Zero/Span Checks for Carbon Monoxide (CO)**

Revision to the warning and control limits for Carbon Monoxide (CO) Zero criteria during Zero/span checks as defined in DAQ-04-001.2 and corresponding elog. This change will be added to SOP DAQ-04-001.2, Section 5.0 during the next SOP review and revision.

Criteria for PZS's (Section 5.0)

	Zero	Span	Precision (Span 1)
Concentration (parts per billion, or ppb)	0	4000	500
Control Limit (±)	< ± 41.0 ppb	< ± 10.1 %	< ± 10.1 %
Warning Limit (±)	< ± 35.0 ppb	± 5%	± 7%

Replace and Discard Original	<input type="checkbox"/>
Add Material to Document	<input type="checkbox"/>
Retain this bulleting until further notice	<input checked="" type="checkbox"/>
Discard this bulleting after noting contents	<input type="checkbox"/>
This bulleting will be invalid after:	Click or tap to enter a date.
This bulletin will be incorporated into DAQ-04-001.2 by	9/30/2022