

DIVISION OF AIR QUALITY
Ambient Monitoring Section
217 West Jones Street / 1641 Mail Service Center
Raleigh, NC 27699-1641

5/4/2022

QA BULLETIN NUMBER DAQ-17-008

TO: Chaitali Bhaumik (LAB Chemist): DocuSigned by: Chaitali Bhaumik 5/4/2022
BE2629DB87740F...

Nathan Miller (LAB Chemistry Tech): DocuSigned by: Nathan Miller 5/5/2022
A575C2DC109A4CA...

THROUGH: Patrick Butler (AMS Chief): DocuSigned by: Patrick Butler 5/4/2022
4290A7E2D2A0429...

Jim Bowyer (LAB Supervisor): DocuSigned by: Jim Bowyer 5/4/2022
C96894826C9B41D...

Joette Steger (PPB Supervisor): DocuSigned by: Joette Steger 5/4/2022
B278F8186CEC4DC...

FROM: Steven Walters (PPB Chemist): DocuSigned by: Steven Walters 5/4/2022
AE6DEE7D0265442...

SUBJECT: ATEC Carbonyl Sampler Certification Criteria Update

ATEC 8000 PAMS Carbonyl Sampler Certification criteria in SOP DAQ-09-002.1 will be changed to all compounds must be ≤ 0.2 micrograms/cartridge after subtracting the method blank value (micrograms/cartridge) from the sampler certification value (micrograms/cartridge).

ATEC 2200 UAT Carbonyl Sampler Certification criteria in SOP DAQ-03-002.1 will change to all compounds must be ≤ 0.2 ppbv after subtracting the method blank value (ppbv) from the sampler certification value (ppbv).

The criteria will be updated in both SOPs during the next revision of the documents.

Replace and Discard Original	<input type="checkbox"/>
Add Material to Document	<input type="checkbox"/>
Retain this bulleting until further notice	<input checked="" type="checkbox"/>
Discard this bulleting after noting contents	<input type="checkbox"/>
This bulleting will be invalid after:	Click or tap to enter a date.
This bulletin will be incorporated into SOP DAQ-09-002.1 and SOP DAQ-03-002.1 by	9/30/2022