

Anti-Static Device Requirement for Total and Total Dissolved Residue Policy
(NC WW/GW LC Branch 6/7/2021)

Plastic type bags have been known to hold a static charge, which may cause weight fluctuations on an analytical balance. When using StableWeigh™, or equivalent bags for Total and/or Total Dissolved Residue, anti-static devices must be in place. This is important for achieving the method specified constant weight requirement of $\pm 0.5\text{mg}$.