

Free Available Chlorine Hold Time Policy (NC WW/GW LC Policy 06/27/2019)

Free Available Chlorine samples must be collected in polyethylene or glass containers and analyzed within 15 minutes of collection. Sample analysis begins when the reagents are added to the sample.